

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
 Data File : VD072934.D
 Acq On : 11 May 2022 16:46
 Operator : VA/SY
 Sample : N2749-04
 Misc : 3.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: May 12 04:13:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	376313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	680352	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	700076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	320459	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	207396	42.457	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.920%
35) Dibromofluoromethane	7.908	113	217025	45.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.840%
50) Toluene-d8	10.332	98	851620	49.687	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.620	95	331190	51.822	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
Data File : VD072934.D
Acq On : 11 May 2022 16:46
Operator : VA/SY
Sample : N2749-04
Misc : 3.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: May 12 04:13:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

