

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
 Data File : VD072937.D
 Acq On : 11 May 2022 18:12
 Operator : VA/SY
 Sample : N2749-13
 Misc : 4.56G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5

Quant Time: May 12 04:14:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	374978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	663239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	651644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	295743	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	208090	42.751	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.500%
35) Dibromofluoromethane	7.908	113	217633	46.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.440%
50) Toluene-d8	10.332	98	808701	48.400	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.800%
62) 4-Bromofluorobenzene	12.620	95	297952	47.824	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
Data File : VD072937.D
Acq On : 11 May 2022 18:12
Operator : VA/SY
Sample : N2749-13
Misc : 4.56G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5

Quant Time: May 12 04:14:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

