

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
 Data File : VD072938.D
 Acq On : 11 May 2022 18:41
 Operator : VA/SY
 Sample : N2749-16
 Misc : 4.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6

Quant Time: May 12 04:15:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.973	168	370577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	661882	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	661001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	311211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	204829	42.581	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.160%
35) Dibromofluoromethane	7.909	113	213918	46.019	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.040%
50) Toluene-d8	10.332	98	808018	48.458	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.920%
62) 4-Bromofluorobenzene	12.620	95	301478	48.489	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.980%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051122\
 Data File : VD072938.D
 Acq On : 11 May 2022 18:41
 Operator : VA/SY
 Sample : N2749-16
 Misc : 4.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6

Quant Time: May 12 04:15:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

