

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051325\
 Data File : VD080303.D
 Acq On : 13 May 2025 15:55
 Operator : RP/MD
 Sample : Q1986-01
 Misc : 5.00G/10ml/MSVOA_D/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-8

Quant Time: May 13 16:08:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	110150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	217017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	185804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	83142	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	72049	50.581	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.160%
35) Dibromofluoromethane	7.805	113	75488	51.899	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.800%
50) Toluene-d8	10.270	98	267257	46.633	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.260%
62) 4-Bromofluorobenzene	12.570	95	71460	38.848	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051325\
Data File : VD080303.D
Acq On : 13 May 2025 15:55
Operator : RP/MD
Sample : Q1986-01
Misc : 5.00G/10ml/MSVOA_D/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-8

Quant Time: May 13 16:08:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
Quant Title : SW846 8260
QLast Update : Wed May 07 08:57:11 2025
Response via : Initial Calibration

