

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051425\
 Data File : VD080328.D
 Acq On : 14 May 2025 17:55
 Operator : RP/MD
 Sample : Q2019-03
 Misc : 4.23G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-K-VOC

Quant Time: May 15 03:38:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	66044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.782	114	174278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	174570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	79539	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.235	65	63143	73.933	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.860%
35) Dibromofluoromethane	7.811	113	71235	60.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.960%
50) Toluene-d8	10.270	98	238714	51.868	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.740%
62) 4-Bromofluorobenzene	12.570	95	65110	44.076	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051425\
Data File : VD080328.D
Acq On : 14 May 2025 17:55
Operator : RP/MD
Sample : Q2019-03
Misc : 4.23G/10ml/MSVOA_D/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-K-VOC

Quant Time: May 15 03:38:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
Quant Title : SW846 8260
QLast Update : Wed May 07 08:57:11 2025
Response via : Initial Calibration

