

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051524\
 Data File : VD078923.D
 Acq On : 15 May 2024 10:37
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 16 01:22:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 14 06:21:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.467	0.464	0.6	119	0.00
3 P	Chloromethane	0.712	0.646	9.3	112	0.00
4 C	Vinyl Chloride	1.077	1.040	3.4#	118	0.00
5 T	Bromomethane	1.041	0.908	12.8	114	0.00
6 T	Chloroethane	0.894	0.833	6.8	111	0.00
7 T	Trichlorofluoromethane	0.979	0.843	13.9	98	0.00
8 T	Diethyl Ether	0.246	0.234	4.9	121	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.517	0.0	117	-0.01
10 T	Methyl Iodide	0.618	0.627	-1.5	111	0.00
11 T	Tert butyl alcohol	0.022	0.021	4.5	117	-0.02
12 CM	1,1-Dichloroethene	0.506	0.506	0.0	119	0.00
13 T	Acrolein	0.047	0.040	14.9	86	0.00
14 T	Allyl chloride	0.494	0.486	1.6	112	0.00
15 T	Acrylonitrile	0.091	0.093	-2.2	120	0.00
16 T	Acetone	0.075	0.082	-9.3	109	-0.01
17 T	Carbon Disulfide	1.628	1.472	9.6	112	0.00
18 T	Methyl Acetate	0.172	0.175	-1.7	121	0.00
19 T	Methyl tert-butyl Ether	0.986	1.013	-2.7	113	0.00
20 T	Methylene Chloride	0.773	0.588	23.9	107	0.00
21 T	trans-1,2-Dichloroethene	0.564	0.574	-1.8	118	-0.01
22 T	Diisopropyl ether	1.117	1.170	-4.7	112	-0.01
23 T	Vinyl Acetate	0.670	0.716	-6.9	116	0.00
24 P	1,1-Dichloroethane	0.886	0.876	1.1	115	0.00
25 T	2-Butanone	0.100	0.108	-8.0	115	0.00
26 T	2,2-Dichloropropane	0.778	0.793	-1.9	121	0.00
27 T	cis-1,2-Dichloroethene	0.662	0.664	-0.3	114	0.00
28 T	Bromochloromethane	0.295	0.287	2.7	105	0.00
29 T	Tetrahydrofuran	0.058	0.062	-6.9	121	0.00
30 C	Chloroform	1.003	0.997	0.6#	116	0.00
31 T	Cyclohexane	0.685	0.653	4.7	114	0.00
32 T	1,1,1-Trichloroethane	0.859	0.878	-2.2	120	0.00
33 S	1,2-Dichloroethane-d4	0.469	0.498	-6.2	119	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.333	0.360	-8.1	115	0.00
36 T	1,1-Dichloropropene	0.414	0.428	-3.4	115	0.00
37 T	Ethyl Acetate	0.141	0.144	-2.1	115	0.00
38 T	Carbon Tetrachloride	0.485	0.507	-4.5	118	0.00
39 T	Methylcyclohexane	0.503	0.523	-4.0	114	0.00
40 TM	Benzene	1.340	1.400	-4.5	116	0.00
41 T	Methacrylonitrile	0.076	0.077	-1.3	122	0.00
42 TM	1,2-Dichloroethane	0.324	0.338	-4.3	118	0.00
43 T	Isopropyl Acetate	0.250	0.267	-6.8	113	0.00
44 TM	Trichloroethene	0.355	0.362	-2.0	117	0.00
45 C	1,2-Dichloropropane	0.294	0.301	-2.4#	114	0.00
46 T	Dibromomethane	0.194	0.198	-2.1	118	0.00
47 T	Bromodichloromethane	0.466	0.482	-3.4	116	0.00
48 T	Methyl methacrylate	0.122	0.134	-9.8	125	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051524\
 Data File : VD078923.D
 Acq On : 15 May 2024 10:37
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: May 16 01:22:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 14 06:21:28 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	111	0.00
50 S	Toluene-d8	1.248	1.406	-12.7	114	0.00
51 T	4-Methyl-2-Pentanone	0.133	0.147	-10.5	117	0.00
52 CM	Toluene	0.858	0.935	-9.0#	120	0.00
53 T	t-1,3-Dichloropropene	0.396	0.422	-6.6	116	0.00
54 T	cis-1,3-Dichloropropene	0.476	0.505	-6.1	114	0.00
55 T	1,1,2-Trichloroethane	0.250	0.268	-7.2	120	0.00
56 T	Ethyl methacrylate	0.258	0.299	-15.9	118	0.00
57 T	1,3-Dichloropropane	0.390	0.414	-6.2	116	0.00
58 T	2-Chloroethyl Vinyl ether	0.111	0.111	0.0	97	0.00
59 T	2-Hexanone	0.093	0.108	-16.1	113	0.00
60 T	Dibromochloromethane	0.341	0.363	-6.5	116	0.00
61 T	1,2-Dibromoethane	0.228	0.245	-7.5	115	0.00
62 S	4-Bromofluorobenzene	0.351	0.404	-15.1	118	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.317	0.340	-7.3	121	0.00
65 PM	Chlorobenzene	1.040	1.094	-5.2	116	0.00
66 T	1,1,1,2-Tetrachloroethane	0.360	0.384	-6.7	120	0.00
67 C	Ethyl Benzene	1.673	1.834	-9.6#	118	0.00
68 T	m/p-Xylenes	0.680	0.775	-14.0	120	0.00
69 T	o-Xylene	0.620	0.681	-9.8	113	0.00
70 T	Styrene	1.099	1.261	-14.7	118	0.00
71 P	Bromoform	0.206	0.225	-9.2	123	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	2.984	3.220	-7.9	119	0.00
74 T	N-amyl acetate	0.484	0.512	-5.8	116	0.00
75 P	1,1,2,2-Tetrachloroethane	0.538	0.565	-5.0	121	0.00
76 T	1,2,3-Trichloropropane	0.352	0.358	-1.7	97	0.00
77 T	Bromobenzene	0.785	0.800	-1.9	117	0.00
78 T	n-propylbenzene	3.607	3.870	-7.3	118	0.00
79 T	2-Chlorotoluene	2.005	2.102	-4.8	118	0.00
80 T	1,3,5-Trimethylbenzene	2.479	2.687	-8.4	119	0.00
81 T	trans-1,4-Dichloro-2-butene	0.178	0.201	-12.9	126	0.00
82 T	4-Chlorotoluene	2.103	2.221	-5.6	117	0.00
83 T	tert-Butylbenzene	2.193	2.406	-9.7	119	0.00
84 T	1,2,4-Trimethylbenzene	2.568	2.781	-8.3	120	0.00
85 T	sec-Butylbenzene	3.271	3.523	-7.7	119	0.00
86 T	p-Isopropyltoluene	2.850	3.123	-9.6	122	0.00
87 T	1,3-Dichlorobenzene	1.661	1.711	-3.0	119	0.00
88 T	1,4-Dichlorobenzene	1.626	1.677	-3.1	119	0.00
89 T	n-Butylbenzene	2.554	2.695	-5.5	119	0.00
90 T	Hexachloroethane	0.601	0.612	-1.8	119	0.00
91 T	1,2-Dichlorobenzene	1.389	1.444	-4.0	117	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.072	0.077	-6.9	116	0.00
93 T	1,2,4-Trichlorobenzene	0.816	0.860	-5.4	120	0.00
94 T	Hexachlorobutadiene	0.426	0.427	-0.2	121	0.00
95 T	Naphthalene	1.390	1.572	-13.1	124	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051524\
Data File : VD078923.D
Acq On : 15 May 2024 10:37
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: May 16 01:22:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
Quant Title : SW846 8260
QLast Update : Tue May 14 06:21:28 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.726	0.771	-6.2	125	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5