



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051524\
Data File : VD078922.D
Acq On : 15 May 2024 09:23
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
Title : SW846 8260
Last Update : Tue May 14 06:21:28 2024



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	7649	PASS
75	95	30	60	48.7	22901	PASS
95	95	100	100	100.0	47069	PASS
96	95	5	9	6.7	3140	PASS
173	174	0.00	2	1.3	608	PASS
174	95	50	100	99.1	46661	PASS
175	174	5	9	6.9	3201	PASS
176	174	95	101	95.2	44429	PASS
177	176	5	9	6.0	2644	PASS