

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051525\
 Data File : VD080344.D
 Acq On : 15 May 2025 13:20
 Operator : RP/MD
 Sample : Q2024-01
 Misc : 5.11G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-02-051325

Quant Time: May 16 02:56:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	95332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.782	114	204434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	195831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	85385	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	65948	53.494	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.980%
35) Dibromofluoromethane	7.805	113	73957	53.976	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.960%
50) Toluene-d8	10.270	98	269779	49.971	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.570	95	76143	43.942	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051525\
Data File : VD080344.D
Acq On : 15 May 2025 13:20
Operator : RP/MD
Sample : Q2024-01
Misc : 5.11G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-02-051325

Quant Time: May 16 02:56:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
Quant Title : SW846 8260
QLast Update : Wed May 07 08:57:11 2025
Response via : Initial Calibration

