

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
 Data File : VD080367.D
 Acq On : 16 May 2025 13:21
 Operator : RP/MD
 Sample : Q2057-03
 Misc : 4.55G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
ClientSampleId :
 MH-L-VOC

Quant Time: May 17 01:19:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.876	168	12698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.782	114	23626	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	15365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	6942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	11341	69.065	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.140%
35) Dibromofluoromethane	7.811	113	10025	63.309	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.620%
50) Toluene-d8	10.270	98	19587	31.393	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	62.780%
62) 4-Bromofluorobenzene	12.576	95	5821	29.067	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	58.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
 Data File : VD080367.D
 Acq On : 16 May 2025 13:21
 Operator : RP/MD
 Sample : Q2057-03
 Misc : 4.55G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-L-VOC

Quant Time: May 17 01:19:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

