

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
 Data File : VD080370.D
 Acq On : 16 May 2025 14:40
 Operator : RP/MD
 Sample : Q2048-11
 Misc : 5.07G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 L2-WC-3-VOC

Quant Time: May 17 01:20:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	71437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.782	114	129699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.576	117	116705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	52253	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.235	65	37106	40.167	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.340%
35) Dibromofluoromethane	7.805	113	46684	53.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.400%
50) Toluene-d8	10.270	98	168921	49.318	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.640%
62) 4-Bromofluorobenzene	12.570	95	44570	40.542	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
Data File : VD080370.D
Acq On : 16 May 2025 14:40
Operator : RP/MD
Sample : Q2048-11
Misc : 5.07G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
L2-WC-3-VOC

Quant Time: May 17 01:20:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
Quant Title : SW846 8260
QLast Update : Wed May 07 08:57:11 2025
Response via : Initial Calibration

