

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
 Data File : VD080372.D
 Acq On : 16 May 2025 15:33
 Operator : RP/MD
 Sample : Q2021-02
 Misc : 5.26G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: May 17 01:20:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 07 08:57:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	34765	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	63747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	60599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	26581	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.235	65	29008	64.524	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.040%
35) Dibromofluoromethane	7.811	113	26200	61.322	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.640%
50) Toluene-d8	10.270	98	81694	48.528	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.060%
62) 4-Bromofluorobenzene	12.575	95	21452	39.702	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051625\
Data File : VD080372.D
Acq On : 16 May 2025 15:33
Operator : RP/MD
Sample : Q2021-02
Misc : 5.26G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: May 17 01:20:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050625S.M
Quant Title : SW846 8260
QLast Update : Wed May 07 08:57:11 2025
Response via : Initial Calibration

