

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
 Data File : VD073059.D
 Acq On : 17 May 2022 16:14
 Operator : VA/SY
 Sample : VD0517SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0517SBL01

Quant Time: May 18 03:20:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 17 13:03:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	341601	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	615603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	619485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	287228	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	192897	45.154	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.300%
35) Dibromofluoromethane	7.908	113	203372	47.053	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.100%
50) Toluene-d8	10.332	98	753788	48.073	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.620	95	278244	48.993	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
Data File : VD073059.D
Acq On : 17 May 2022 16:14
Operator : VA/SY
Sample : VD0517SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0517SBL01

Quant Time: May 18 03:20:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Tue May 17 13:03:08 2022
Response via : Initial Calibration

