

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
 Data File : VD073064.D
 Acq On : 17 May 2022 18:37
 Operator : VA/SY
 Sample : N2855-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-051322

Quant Time: May 18 05:09:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	79011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	145262	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	160885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	78107	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	72064	72.932	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.860%
35) Dibromofluoromethane	7.909	113	59731	58.565	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.140%
50) Toluene-d8	10.332	98	180554	48.798	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.600%
62) 4-Bromofluorobenzene	12.620	95	69319	51.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
Data File : VD073064.D
Acq On : 17 May 2022 18:37
Operator : VA/SY
Sample : N2855-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-051322

Quant Time: May 18 05:09:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

