

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
 Data File : VD073066.D
 Acq On : 17 May 2022 19:35
 Operator : VA/SY
 Sample : N2847-01
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 REDBOX-164718

Quant Time: May 18 05:10:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	237040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	434118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	453888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	212614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	153123	51.655	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.300%	
35) Dibromofluoromethane	7.908	113	151364	49.660	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.320%	
50) Toluene-d8	10.332	98	544992	49.287	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.580%	
62) 4-Bromofluorobenzene	12.620	95	197313	49.267	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.540%	
Target Compounds						Qvalue
16) Acetone	4.156	43	24157	45.032	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
Data File : VD073066.D
Acq On : 17 May 2022 19:35
Operator : VA/SY
Sample : N2847-01
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
REDBOX-164718

Quant Time: May 18 05:10:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

