

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
 Data File : VD073080.D
 Acq On : 18 May 2022 03:14
 Operator : VA/SY
 Sample : N2879-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1-0516

Quant Time: May 18 05:14:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 18 04:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	303968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	554379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	566983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	260300	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	171558	45.131	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.260%
35) Dibromofluoromethane	7.908	113	186497	47.914	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.820%
50) Toluene-d8	10.332	98	687239	48.669	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.620	95	249064	48.698	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
Data File : VD073080.D
Acq On : 18 May 2022 03:14
Operator : VA/SY
Sample : N2879-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1-0516

Quant Time: May 18 05:14:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
Quant Title : SW846 8260
QLast Update : Wed May 18 04:58:34 2022
Response via : Initial Calibration

