



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD051722\
 Data File : VD073069.D
 Acq On : 17 May 2022 21:01
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051722S.M
 Title : SW846 8260
 Last Update : Wed May 18 04:58:34 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	25528	PASS
75	95	30	60	51.5	65285	PASS
95	95	100	100	100.0	126832	PASS
96	95	5	9	6.8	8578	PASS
173	174	0.00	2	1.5	1519	PASS
174	95	50	100	77.9	98789	PASS
175	174	5	9	7.4	7261	PASS
176	174	95	101	95.9	94720	PASS
177	176	5	9	6.8	6474	PASS