

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073231.D
 Acq On : 21 May 2022 15:55
 Operator : VA/SY
 Sample : N2870-02
 Misc : 4.90G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P033-SS001-0612-01

Quant Time: May 22 01:32:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	308389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	544674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	561356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	252038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	168159	49.340	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.680%
35) Dibromofluoromethane	7.909	113	180369	50.184	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.360%
50) Toluene-d8	10.332	98	669035	50.498	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.000%
62) 4-Bromofluorobenzene	12.620	95	243188	52.303	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.600%
Target Compounds						
2) Dichlorodifluoromethane	1.991	85	5698	1.849	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073231.D
Acq On : 21 May 2022 15:55
Operator : VA/SY
Sample : N2870-02
Misc : 4.90G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P033-SS001-0612-01

Quant Time: May 22 01:32:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

