

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073232.D
 Acq On : 21 May 2022 16:25
 Operator : VA/SY
 Sample : N2870-03
 Misc : 5.57G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P033-SS001-1218-01

Quant Time: May 22 01:32:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	301298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	534887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	570097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	263716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	176761	53.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.180%	
35) Dibromofluoromethane	7.903	113	182383	51.673	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.340%	
50) Toluene-d8	10.332	98	662094	50.888	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.780%	
62) 4-Bromofluorobenzene	12.620	95	245405	53.746	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.500%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.167	59	16758	32.438	ug/l #	75
16) Acetone	4.162	43	11978	20.768	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073232.D
Acq On : 21 May 2022 16:25
Operator : VA/SY
Sample : N2870-03
Misc : 5.57G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P033-SS001-1218-01

Quant Time: May 22 01:32:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

