

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073234.D
 Acq On : 21 May 2022 17:23
 Operator : VA/SY
 Sample : N2870-05
 Misc : 5.47G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P033-SS001-2430-01

Quant Time: May 22 01:32:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	294989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	526355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	571834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	263917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	173662	53.270	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.540%	
35) Dibromofluoromethane	7.909	113	184904	53.237	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.480%	
50) Toluene-d8	10.332	98	668962	52.250	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.500%	
62) 4-Bromofluorobenzene	12.620	95	246374	54.833	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.660%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	6098	2.069	ug/l	97
16) Acetone	4.156	43	6952	12.312	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073234.D
Acq On : 21 May 2022 17:23
Operator : VA/SY
Sample : N2870-05
Misc : 5.47G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P033-SS001-2430-01

Quant Time: May 22 01:32:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

