

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073236.D
 Acq On : 21 May 2022 18:22
 Operator : VA/SY
 Sample : N2870-07
 Misc : 6.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P034-SS001-0612-01

Quant Time: May 22 01:33:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	293941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	520292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	548857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	250243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	168737	51.944	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.880%	
35) Dibromofluoromethane	7.908	113	176807	51.499	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.000%	
50) Toluene-d8	10.332	98	650706	51.416	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.840%	
62) 4-Bromofluorobenzene	12.620	95	237559	53.487	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.980%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	4913	1.673	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073236.D
Acq On : 21 May 2022 18:22
Operator : VA/SY
Sample : N2870-07
Misc : 6.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P034-SS001-0612-01

Quant Time: May 22 01:33:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

