

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073240.D
 Acq On : 21 May 2022 20:21
 Operator : VA/SY
 Sample : N2920-01
 Misc : 4.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P015-SS006-0006-01

Quant Time: May 22 01:34:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	291182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	512312	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	545814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	237032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	162842	50.604	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.200%	
35) Dibromofluoromethane	7.908	113	174922	51.743	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.480%	
50) Toluene-d8	10.332	98	645512	51.800	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.600%	
62) 4-Bromofluorobenzene	12.620	95	234614	53.647	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.300%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.985	85	5644	1.940	ug/l	99
70) Styrene	12.185	104	38315	3.174	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073240.D
Acq On : 21 May 2022 20:21
Operator : VA/SY
Sample : N2920-01
Misc : 4.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P015-SS006-0006-01

Quant Time: May 22 01:34:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

