

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073242.D
 Acq On : 21 May 2022 21:20
 Operator : VA/SY
 Sample : N2920-03
 Misc : 5.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P015-SS008-0006-01

Quant Time: May 22 01:34:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	286914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	494998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	521822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	247864	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	156221	49.269	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.540%
35) Dibromofluoromethane	7.908	113	171160	52.401	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.800%
50) Toluene-d8	10.332	98	623069	51.748	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.620	95	229755	54.373	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073242.D
Acq On : 21 May 2022 21:20
Operator : VA/SY
Sample : N2920-03
Misc : 5.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P015-SS008-0006-01

Quant Time: May 22 01:34:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

