

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073244.D
 Acq On : 21 May 2022 22:19
 Operator : VA/SY
 Sample : N2920-05
 Misc : 4.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P015-SS010-0006-01

Quant Time: May 22 01:35:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	271033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	481582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	488111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	201560	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	148336	49.523	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.040%
35) Dibromofluoromethane	7.909	113	164141	51.652	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.300%
50) Toluene-d8	10.332	98	593365	50.654	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.300%
62) 4-Bromofluorobenzene	12.620	95	200143	48.685	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073244.D
Acq On : 21 May 2022 22:19
Operator : VA/SY
Sample : N2920-05
Misc : 4.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P015-SS010-0006-01

Quant Time: May 22 01:35:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

