

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073257.D
 Acq On : 22 May 2022 06:03
 Operator : VA/SY
 Sample : N2920-10
 Misc : 5.41G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P024-SS001-2430-01

Quant Time: May 23 01:13:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	272116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	488717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	509925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	229015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	156729	52.117	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.240%	
35) Dibromofluoromethane	7.903	113	166815	51.728	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.460%	
50) Toluene-d8	10.332	98	598401	50.338	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	12.620	95	220451	52.842	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.680%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.997	85	3598	1.323	ug/l	82
16) Acetone	4.162	43	7916	15.197	ug/l #	84
31) Cyclohexane	7.979	56	7333	1.395	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073257.D
Acq On : 22 May 2022 06:03
Operator : VA/SY
Sample : N2920-10
Misc : 5.41G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P024-SS001-2430-01

Quant Time: May 23 01:13:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

