

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073260.D
 Acq On : 22 May 2022 07:31
 Operator : VA/SY
 Sample : N2920-15
 Misc : 5.26G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P024-SS003-0006-01

Quant Time: May 23 01:13:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	290814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	520921	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	537577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	228536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	157191	48.910	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.820%
35) Dibromofluoromethane	7.908	113	172990	50.326	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.660%
50) Toluene-d8	10.332	98	640261	50.529	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.060%
62) 4-Bromofluorobenzene	12.620	95	221123	49.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.460%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	2949	1.015	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073260.D
Acq On : 22 May 2022 07:31
Operator : VA/SY
Sample : N2920-15
Misc : 5.26G/5.00ml/MSVOA_D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P024-SS003-0006-01

Quant Time: May 23 01:13:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

