

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073269.D
 Acq On : 22 May 2022 11:53
 Operator : VA/SY
 Sample : N2920-24
 Misc : 4.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P028-SS001-1218-01

Quant Time: May 23 01:15:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	309968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	564425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	621471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	311847	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	220182	64.276	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	128.560%	
35) Dibromofluoromethane	7.903	113	200380	53.801	ug/l	-0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	107.600%	
50) Toluene-d8	10.332	98	710909	51.781	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	103.560%	
62) 4-Bromofluorobenzene	12.620	95	298604	61.975	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	123.940%	
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	12843	21.645	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073269.D
Acq On : 22 May 2022 11:53
Operator : VA/SY
Sample : N2920-24
Misc : 4.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P028-SS001-1218-01

Quant Time: May 23 01:15:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

