

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073235.D
 Acq On : 21 May 2022 17:53
 Operator : VA/SY
 Sample : N2870-06
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P034-SS001-0006-01

Quant Time: May 22 01:33:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	292702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	520257	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	541093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	229952	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	171310	52.959	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.920%
35) Dibromofluoromethane	7.908	113	180123	52.468	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.940%
50) Toluene-d8	10.332	98	638047	50.419	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.840%
62) 4-Bromofluorobenzene	12.620	95	228435	51.436	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.880%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073235.D
Acq On : 21 May 2022 17:53
Operator : VA/SY
Sample : N2870-06
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P034-SS001-0006-01

Quant Time: May 22 01:33:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. 0.000 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53




Instrument :
MSVOA_D
ClientSampleId :
P034-SS001-0006-01


Tgt Ion:168 Resp: 292702
Ion Ratio Lower Upper
168 100
99 58.3 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 52.959 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53

Tgt Ion: 65 Resp: 171310
Ion Ratio Lower Upper
65 100
67 54.5 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.861 min Scan# 1179

Delta R.T. 0.006 min

Lab File: VD073235.D

Acq: 21 May 2022 17:53

Instrument :

MSVOA_D

ClientSampleId :

P034-SS001-0006-01





Tgt Ion:114 Resp: 520257

Ion Ratio Lower Upper

114 100

63 20.0 0.0 46.6

88 14.8 0.0 33.2



#35

Dibromofluoromethane

Concen: 52.468 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. -0.006 min

Lab File: VD073235.D

Acq: 21 May 2022 17:53




Tgt Ion:113 Resp: 180123


Ion Ratio Lower Upper

113 100

111 103.2 83.8 125.6

192 18.1 14.6 21.8









#50
Toluene-d8
Concen: 50.419 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53




Instrument :
MSVOA_D
ClientSampleId :
P034-SS001-0006-01


Tgt Ion: 98 Resp: 638047
Ion Ratio Lower Upper
98 100
100 66.6 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 51.436 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53

Tgt Ion: 95 Resp: 228435
Ion Ratio Lower Upper
95 100
174 77.5 0.0 152.8
176 75.6 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 117
Delta R.T. -0.006 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53

Instrument :
MSVOA_D
ClientSampleId :
P034-SS001-0006-01

Tgt Ion:117	Resp: 541093
Ion Ratio	Lower Upper
117 100	
82 55.0	45.3 67.9
119 32.6	26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD073235.D
Acq: 21 May 2022 17:53

Tgt Ion:152	Resp: 229952
Ion Ratio	Lower Upper
152 100	
115 61.8	31.3 93.8
150 160.1	0.0 348.6

