

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073265.D
 Acq On : 22 May 2022 09:56
 Operator : VA/SY
 Sample : N2920-20
 Misc : 5.38G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P024-SS008-0006-01

Quant Time: May 23 01:15:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	316916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	578405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	627234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	314143	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	222496	63.527	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.060%
35) Dibromofluoromethane	7.908	113	211184	55.332	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.660%
50) Toluene-d8	10.332	98	722515	51.354	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.700%
62) 4-Bromofluorobenzene	12.620	95	291759	59.090	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.180%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
Data File : VD073265.D
Acq On : 22 May 2022 09:56
Operator : VA/SY
Sample : N2920-20
Misc : 5.38G/5.00ml/MSVOA_D/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P024-SS008-0006-01

Quant Time: May 23 01:15:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.973 min Scan# 1029

Delta R.T. -0.000 min

Lab File: VD073265.D


Acq: 22 May 2022 09:56

Instrument :

MSVOA_D

ClientSampleId :

P024-SS008-0006-01






Tgt Ion:168 Resp: 316916

Ion Ratio Lower Upper

168 100

99 60.3 49.8 74.6



1,2-Dichloroethane-d4

Concen: 63.527 ug/l

RT: 8.326 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VD073265.D




Acq: 22 May 2022 09:56


Tgt Ion: 65 Resp: 222496

Ion Ratio Lower Upper

65 100

67 52.3 0.0 101.4








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.855 min Scan# 1179
Delta R.T. 0.000 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56




Instrument :
MSVOA_D
ClientSampleId :
P024-SS008-0006-01


Tgt Ion:114 Resp: 578405
Ion Ratio Lower Upper
114 100
63 19.7 0.0 46.6
88 16.1 0.0 33.2



#35
Dibromofluoromethane
Concen: 55.332 ug/l
RT: 7.908 min Scan# 1018
Delta R.T. -0.006 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56

Tgt Ion:113 Resp: 211184
Ion Ratio Lower Upper
113 100
111 100.4 83.8 125.6
192 17.7 14.6 21.8









#50
Toluene-d8
Concen: 51.354 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56




Instrument :
MSVOA_D
ClientSampleId :
P024-SS008-0006-01


Tgt Ion: 98 Resp: 722515
Ion Ratio Lower Upper
98 100
100 66.2 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 59.090 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56

Tgt Ion: 95 Resp: 291759
Ion Ratio Lower Upper
95 100
174 77.1 0.0 152.8
176 74.0 0.0 145.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.637 min Scan# 117
Delta R.T. -0.001 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56

Instrument :
MSVOA_D
ClientSampleId :
P024-SS008-0006-01

Tgt Ion:117 Resp: 627234
Ion Ratio Lower Upper
117 100
82 55.9 45.3 67.9
119 31.8 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD073265.D
Acq: 22 May 2022 09:56

Tgt Ion:152 Resp: 314143
Ion Ratio Lower Upper
152 100
115 61.2 31.3 93.8
150 159.8 0.0 348.6

