



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052122\
 Data File : VD073226.D
 Acq On : 21 May 2022 13:02
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Title : SW846 8260
 Last Update : Thu May 19 15:13:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	26669	PASS
75	95	30	60	51.0	75331	PASS
95	95	100	100	100.0	147840	PASS
96	95	5	9	6.5	9594	PASS
173	174	0.00	2	1.8	1993	PASS
174	95	50	100	76.5	113099	PASS
175	174	5	9	7.8	8792	PASS
176	174	95	101	96.8	109440	PASS
177	176	5	9	6.3	6939	PASS