

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052224\
 Data File : VD078981.D
 Acq On : 22 May 2024 15:59
 Operator : RP/MD
 Sample : P2525-03
 Misc : 4.57G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC-MH-AA

Quant Time: May 23 06:05:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 14 06:21:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.875	168	119271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	247794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	237847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	106814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	64054	57.254	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.500%
35) Dibromofluoromethane	7.804	113	83441	50.586	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.180%
50) Toluene-d8	10.269	98	300961	48.670	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.340%
62) 4-Bromofluorobenzene	12.569	95	82080	47.250	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.500%
Target Compounds						
16) Acetone	4.022	43	3264	18.298	ug/l	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052224\
Data File : VD078981.D
Acq On : 22 May 2024 15:59
Operator : RP/MD
Sample : P2525-03
Misc : 4.57G/5ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC-MH-AA

Quant Time: May 23 06:05:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051324S.M
Quant Title : SW846 8260
QLast Update : Tue May 14 06:21:28 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59




Instrument :
MSVOA_D
ClientSampleId :
VOC-MH-AA


Tgt Ion:168 Resp: 119271
Ion Ratio Lower Upper
168 100
99 51.8 41.1 61.7



#16
Acetone
Concen: 18.298 ug/l
RT: 4.022 min Scan# 414
Delta R.T. -0.006 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59

Tgt Ion: 43 Resp: 3264
Ion Ratio Lower Upper
43 100
58 37.0 30.4 45.6





#33

1,2-Dichloroethane-d4

Concen: 57.254 ug/l

RT: 8.234 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VD078981.D

Acq: 22 May 2024 15:59

Instrument :

MSVOA_D

ClientSampleId :





VOC-MH-AA

Tgt Ion: 65 Resp: 64054

Ion Ratio Lower Upper

65 100

67 56.6 0.0 111.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.006 min

Lab File: VD078981.D

Acq: 22 May 2024 15:59




Tgt Ion: 114 Resp: 247794


Ion Ratio Lower Upper

114 100

63 15.5 0.0 34.0

88 12.9 0.0 27.2









#35
Dibromofluoromethane
Concen: 50.586 ug/l
RT: 7.804 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59




Instrument :
MSVOA_D
ClientSampleId :
VOC-MH-AA


Tgt Ion:113 Resp: 83441
Ion Ratio Lower Upper
113 100
111 101.2 84.0 126.0
192 19.0 15.4 23.0



#50
Toluene-d8
Concen: 48.670 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59

Tgt Ion: 98 Resp: 300961
Ion Ratio Lower Upper
98 100
100 63.7 51.6 77.4









#62
4-Bromofluorobenzene
Concen: 47.250 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.006 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59




Instrument :
MSVOA_D
ClientSampleId :
VOC-MH-AA


Tgt Ion: 95 Resp: 82080
Ion Ratio Lower Upper
95 100
174 95.0 0.0 195.4
176 92.8 0.0 192.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VD078981.D
Acq: 22 May 2024 15:59

Tgt Ion: 117 Resp: 237847
Ion Ratio Lower Upper
117 100
82 50.4 39.8 59.6
119 31.7 25.7 38.5





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.516 min Scan# 20

Delta R.T. 0.000 min

Lab File: VD078981.D


Acq: 22 May 2024 15:59

Instrument :

MSVOA_D

ClientSampleId :

VOC-MH-AA



Tgt Ion:152 Resp: 106814

Ion Ratio Lower Upper

152 100

115 53.6 26.5 79.5

150 153.0 0.0 341.6

