

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073288.D
 Acq On : 23 May 2022 17:43
 Operator : VA/SY
 Sample : N2979-01
 Misc : 4.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-7-052022

Quant Time: May 24 06:17:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	178900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	315854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	353687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	166461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	123923	62.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.360%
35) Dibromofluoromethane	7.908	113	120083	57.615	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.240%
50) Toluene-d8	10.332	98	397422	51.728	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.460%
62) 4-Bromofluorobenzene	12.620	95	155012	57.492	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.980%
Target Compounds						
16) Acetone	4.150	43	4674	13.649	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.614	75	75821	111.447	ug/l #	7

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073288.D
Acq On : 23 May 2022 17:43
Operator : VA/SY
Sample : N2979-01
Misc : 4.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-7-052022

Quant Time: May 24 06:17:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

