

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073292.D
 Acq On : 23 May 2022 19:41
 Operator : VA/SY
 Sample : N2885-18
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P014-SS002-0612-01

Quant Time: May 24 03:29:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	312679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	559990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	573310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	264220	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	165722	47.958	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.920%
35) Dibromofluoromethane	7.909	113	181881	49.221	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.440%
50) Toluene-d8	10.332	98	682163	50.080	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.620	95	269201	56.315	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.620%
Target Compounds						
16) Acetone	4.156	43	17126	28.613	ug/l	96
52) Toluene	10.391	92	15756	1.557	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073292.D
Acq On : 23 May 2022 19:41
Operator : VA/SY
Sample : N2885-18
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P014-SS002-0612-01

Quant Time: May 24 03:29:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

