

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073296.D
 Acq On : 23 May 2022 22:34
 Operator : VA/SY
 Sample : BLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0523SBL02

Quant Time: May 24 03:30:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	2652	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.856	114	3570	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	3287	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	2549	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	517	17.640	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	35.280%#
35) Dibromofluoromethane	7.914	113	853	36.210	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	72.420%
50) Toluene-d8	10.332	98	4097	47.180	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.360%
62) 4-Bromofluorobenzene	12.614	95	1507	49.450	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073296.D
Acq On : 23 May 2022 22:34
Operator : VA/SY
Sample : BLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0523SBL02

Quant Time: May 24 03:30:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

