

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073299.D
 Acq On : 24 May 2022 00:01
 Operator : VA/SY
 Sample : N2885-19
 Misc : 4.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P014-SS002-1218-01

Quant Time: May 24 03:31:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	311600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	542640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	562841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	258320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	161475	46.891	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.780%
35) Dibromofluoromethane	7.908	113	180292	50.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.700%
50) Toluene-d8	10.332	98	666809	50.518	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.040%
62) 4-Bromofluorobenzene	12.620	95	250049	53.981	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.960%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.967	56	9379	1.559	ug/l #	15

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073299.D
Acq On : 24 May 2022 00:01
Operator : VA/SY
Sample : N2885-19
Misc : 4.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P014-SS002-1218-01

Quant Time: May 24 03:31:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

