

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073304.D
 Acq On : 24 May 2022 02:26
 Operator : VA/SY
 Sample : N2949-04
 Misc : 6.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-0612-01

Quant Time: May 24 03:32:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	260808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	478363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	509377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	231666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	162093	56.237	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 112.480%	
35) Dibromofluoromethane	7.909	113	167099	52.937	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.880%	
50) Toluene-d8	10.332	98	598324	51.421	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.840%	
62) 4-Bromofluorobenzene	12.620	95	227949	55.822	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.640%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.179	59	16768	37.496	ug/l	# 75
16) Acetone	4.156	43	17079	34.210	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073304.D
Acq On : 24 May 2022 02:26
Operator : VA/SY
Sample : N2949-04
Misc : 6.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-0612-01

Quant Time: May 24 03:32:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

