

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073307.D
 Acq On : 24 May 2022 03:52
 Operator : VA/SY
 Sample : N2949-09
 Misc : 4.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-1824-01

Quant Time: May 24 08:16:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	4138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	8264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	8467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	3241	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	4653	101.748	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	203.500%#
35) Dibromofluoromethane	7.903	113	3954	72.509	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	145.020%
50) Toluene-d8	10.326	98	8596	42.763	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.520%
62) 4-Bromofluorobenzene	12.626	95	3432	48.650	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073307.D
Acq On : 24 May 2022 03:52
Operator : VA/SY
Sample : N2949-09
Misc : 4.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-1824-01

Quant Time: May 24 08:16:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

