

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
 Data File : VD073308.D
 Acq On : 24 May 2022 04:21
 Operator : VA/SY
 Sample : N2949-10
 Misc : 6.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-2430-01

Quant Time: May 24 08:16:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	106536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	188817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	225517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	104291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	92472	78.541	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.080%
35) Dibromofluoromethane	7.909	113	83593	67.092	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	134.180%
50) Toluene-d8	10.332	98	239721	52.194	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.380%
62) 4-Bromofluorobenzene	12.620	95	95630	59.331	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.660%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	13114	64.305	ug/l	90
70) Styrene	12.179	104	11244	2.255	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052322\
Data File : VD073308.D
Acq On : 24 May 2022 04:21
Operator : VA/SY
Sample : N2949-10
Misc : 6.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-2430-01

Quant Time: May 24 08:16:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

