

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073318.D
 Acq On : 24 May 2022 12:03
 Operator : VA/SY
 Sample : VD0524SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0524SBL01

Quant Time: May 24 16:43:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	332920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	604671	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	617834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	284776	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	189747	51.572	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.140%
35) Dibromofluoromethane	7.908	113	198800	49.824	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.640%
50) Toluene-d8	10.332	98	734899	49.965	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.620	95	275265	53.328	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073318.D
Acq On : 24 May 2022 12:03
Operator : VA/SY
Sample : VD0524SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0524SBL01

Quant Time: May 24 16:43:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

