

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073324.D
 Acq On : 24 May 2022 14:54
 Operator : VA/SY
 Sample : N2949-07RE
 Misc : 5.13G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-0612-02

Quant Time: May 25 05:26:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	306868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	548176	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	571892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	256354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	166033	48.958	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.920%
35) Dibromofluoromethane	7.908	113	184678	51.055	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.120%
50) Toluene-d8	10.332	98	675964	50.695	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.620	95	247276	52.843	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.680%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073324.D
Acq On : 24 May 2022 14:54
Operator : VA/SY
Sample : N2949-07RE
Misc : 5.13G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-0612-02

Quant Time: May 25 05:26:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

