

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073327.D
 Acq On : 24 May 2022 16:20
 Operator : VA/SY
 Sample : N2949-10
 Misc : 6.14G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-2430-01

Quant Time: May 25 05:27:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	308095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	548999	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	571723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	258717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	174990	51.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.780%	
35) Dibromofluoromethane	7.908	113	186052	51.358	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.720%	
50) Toluene-d8	10.332	98	671363	50.274	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.540%	
62) 4-Bromofluorobenzene	12.620	95	257302	54.903	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.800%	
Target Compounds						
					Qvalue	
70) Styrene	12.185	104	16514	1.306	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073327.D
Acq On : 24 May 2022 16:20
Operator : VA/SY
Sample : N2949-10
Misc : 6.14G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-2430-01

Quant Time: May 25 05:27:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

