

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073345.D
 Acq On : 25 May 2022 02:20
 Operator : VA/SY
 Sample : N2886-01
 Misc : 6.64G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P014-SS002-2430-01

Quant Time: May 25 05:33:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	4664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.850	114	7221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	7556	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	3082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	3680	71.396	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 142.800%	
35) Dibromofluoromethane	7.908	113	3360	70.516	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 141.040%	
50) Toluene-d8	10.332	98	8595	48.934	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 97.860%	
62) 4-Bromofluorobenzene	12.620	95	3282	53.243	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073345.D
Acq On : 25 May 2022 02:20
Operator : VA/SY
Sample : N2886-01
Misc : 6.64G/5.00ml/MSVOA_D/SOIL
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P014-SS002-2430-01

Quant Time: May 25 05:33:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

