

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073347.D
 Acq On : 25 May 2022 03:18
 Operator : VA/SY
 Sample : N2886-03
 Misc : 4.71G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P020-SS001-0006-01

Quant Time: May 25 05:33:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	182165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	325402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	358439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	159628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	127297	63.232	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.460%	
35) Dibromofluoromethane	7.909	113	124156	57.822	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.640%	
50) Toluene-d8	10.332	98	400484	50.597	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.200%	
62) 4-Bromofluorobenzene	12.620	95	150871	54.314	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.620%	
Target Compounds						
16) Acetone	4.168	43	9060	25.982	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073347.D
Acq On : 25 May 2022 03:18
Operator : VA/SY
Sample : N2886-03
Misc : 4.71G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P020-SS001-0006-01

Quant Time: May 25 05:33:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration

