

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
 Data File : VD073325.D
 Acq On : 24 May 2022 15:22
 Operator : VA/SY
 Sample : N2949-08
 Misc : 6.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P043-SS001-1218-01

Quant Time: May 25 05:26:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 19 15:13:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	218043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	392067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	435327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	198482	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	143455	59.533	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.060%
35) Dibromofluoromethane	7.909	113	141787	54.805	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.620%
50) Toluene-d8	10.332	98	482825	50.628	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.620	95	190284	56.855	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073325.D
Acq On : 24 May 2022 15:22
Operator : VA/SY
Sample : N2949-08
Misc : 6.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-1218-01

Quant Time: May 25 05:26:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Quant Title : SW846 8260
QLast Update : Thu May 19 15:13:42 2022
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.973 min Scan# 1029

Delta R.T. 0.000 min

Lab File: VD073325.D


Acq: 24 May 2022 15:22

Instrument :

MSVOA_D

ClientSampleId :

P043-SS001-1218-01






Tgt Ion:168 Resp: 218043

Ion Ratio Lower Upper

168 100

99 64.2 49.8 74.6



1,2-Dichloroethane-d4

Concen: 59.533 ug/l

RT: 8.326 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VD073325.D




Acq: 24 May 2022 15:22


Tgt Ion: 65 Resp: 143455

Ion Ratio Lower Upper

65 100

67 52.9 0.0 101.4








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1179
Delta R.T. 0.006 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22




Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-1218-01

Tgt Ion:114 Resp: 392067
Ion Ratio Lower Upper
114 100
63 19.2 0.0 46.6
88 15.4 0.0 33.2



#35
Dibromofluoromethane
Concen: 54.805 ug/l
RT: 7.909 min Scan# 1018
Delta R.T. -0.005 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22

Tgt Ion:113 Resp: 141787
Ion Ratio Lower Upper
113 100
111 104.1 83.8 125.6
192 18.6 14.6 21.8








#50
Toluene-d8
Concen: 50.628 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22




Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-1218-01


Tgt Ion: 98 Resp: 482825
Ion Ratio Lower Upper
98 100
100 67.0 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 56.855 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. 0.000 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22

Tgt Ion: 95 Resp: 190284
Ion Ratio Lower Upper
95 100
174 76.7 0.0 152.8
176 74.2 0.0 145.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1
Delta R.T. -0.000 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22

Instrument :
MSVOA_D
ClientSampleId :
P043-SS001-1218-01

Tgt Ion:117 Resp: 435327
Ion Ratio Lower Upper
117 100
82 53.0 45.3 67.9
119 32.4 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD073325.D
Acq: 24 May 2022 15:22

Tgt Ion:152 Resp: 198482
Ion Ratio Lower Upper
152 100
115 62.0 31.3 93.8
150 158.7 0.0 348.6

