


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD052422\
Data File : VD073316.D
Acq On : 24 May 2022 09:14
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D051922S.M
Title : SW846 8260
Last Update : Thu May 19 15:13:42 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	25533	PASS
75	95	30	60	50.0	69539	PASS
95	95	100	100	100.0	139147	PASS
96	95	5	9	6.6	9119	PASS
173	174	0.00	2	1.7	1785	PASS
174	95	50	100	76.5	106467	PASS
175	174	5	9	7.1	7576	PASS
176	174	95	101	97.8	104155	PASS
177	176	5	9	6.8	7114	PASS