

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053024\
 Data File : VD079103.D
 Acq On : 30 May 2024 18:23
 Operator : RP/MD
 Sample : P2633-08
 Misc : 5.10G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VHBLK001

Quant Time: May 31 01:59:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 31 01:52:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	215559	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	215014	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	88946	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	160972	24.267	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.080%
7) Chloroethane-d5	2.799	69	141110	22.000	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	3.916	65	24239	20.584	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.320%
21) 2-Butanone-d5	6.987	46	23427	46.174	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.340%
24) Chloroform-d	7.563	84	126902	21.055	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.240%
26) 1,2-Dichloroethane-d4	8.228	65	64046	23.035	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.120%
32) Benzene-d6	8.199	84	260614	21.865	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.440%
36) 1,2-Dichloropropane-d6	9.204	67	74726	21.764	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.040%
41) Toluene-d8	10.269	98	232754	20.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.440%
43) trans-1,3-Dichloroprop...	10.522	79	31141	21.578	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.320%
47) 2-Hexanone-d5	10.875	63	19428	43.769	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.540%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	64370	21.166	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.680%
66) 1,2-Dichlorobenzene-d4	13.810	152	74166	24.161	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.640%
Target Compounds						
16) Methylene chloride	4.793	84	7926	2.064	ug/L	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053024\
Data File : VD079103.D
Acq On : 30 May 2024 18:23
Operator : RP/MD
Sample : P2633-08
Misc : 5.10G/10ml/MSVOA_D/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VHBLK001

Quant Time: May 31 01:59:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 31 01:52:18 2024
Response via : Initial Calibration

