

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053025\
 Data File : VD080400.D
 Acq On : 30 May 2025 11:35
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: May 31 00:44:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D053025S.M
 Quant Title : SW846 8260
 QLast Update : Sat May 31 00:36:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	251065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	603831	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	598174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	242402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	214191	68.328	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.660%
35) Dibromofluoromethane	7.805	113	206212	52.542	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.080%
50) Toluene-d8	10.270	98	864547	54.749	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.570	95	215847	44.655	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD053025\
Data File : VD080400.D
Acq On : 30 May 2025 11:35
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: May 31 00:44:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D053025S.M
Quant Title : SW846 8260
QLast Update : Sat May 31 00:36:06 2025
Response via : Initial Calibration

