

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
 Data File : VD073526.D
 Acq On : 10 Jun 2022 14:07
 Operator : VA/SY
 Sample : N3234-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RB-21499

Quant Time: Jun 10 23:05:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	271688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	520116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	614572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	333988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	134909	46.272	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.540%
35) Dibromofluoromethane	7.908	113	165893	49.601	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.200%
50) Toluene-d8	10.332	98	649061	51.328	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.660%
62) 4-Bromofluorobenzene	12.620	95	304151	66.255	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	132.520%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	20937	40.784	ug/l #	80
39) Methylcyclohexane	9.350	83	9013	1.540	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
Data File : VD073526.D
Acq On : 10 Jun 2022 14:07
Operator : VA/SY
Sample : N3234-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RB-21499

Quant Time: Jun 10 23:05:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

