

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
 Data File : VD073527.D
 Acq On : 10 Jun 2022 14:36
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jun 10 23:05:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	284567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	535617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	603992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	312780	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	145559	47.665	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.340%
35) Dibromofluoromethane	7.908	113	171902	49.910	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.820%
50) Toluene-d8	10.332	98	667209	51.236	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.480%
62) 4-Bromofluorobenzene	12.620	95	288926	61.117	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.240%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	7510	13.967	ug/l #	57
95) Naphthalene	15.379	128	31687	6.773	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
Data File : VD073527.D
Acq On : 10 Jun 2022 14:36
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jun 10 23:05:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

