

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
 Data File : VD073530.D
 Acq On : 10 Jun 2022 16:03
 Operator : VA/SY
 Sample : N3299-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COM-4

Quant Time: Jun 10 23:06:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 12:33:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	249922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	461461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	537925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	276155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	133203	49.666	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.340%
35) Dibromofluoromethane	7.903	113	64027	21.577	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	43.160%#
50) Toluene-d8	10.332	98	587626	52.376	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.760%
62) 4-Bromofluorobenzene	12.620	95	260897	64.057	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.120%
Target Compounds						
					Qvalue	
16) Acetone	4.162	43	7329	15.520	ug/l	93
95) Naphthalene	15.379	128	10008	5.123	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061022\
Data File : VD073530.D
Acq On : 10 Jun 2022 16:03
Operator : VA/SY
Sample : N3299-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COM-4

Quant Time: Jun 10 23:06:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D060922S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 12:33:32 2022
Response via : Initial Calibration

