

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061024\
 Data File : VD079156.D
 Acq On : 10 Jun 2024 14:03
 Operator : RP/MD
 Sample : VIBLK806
 Misc : 5.00G/10ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK806

Quant Time: Jun 11 02:48:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jun 11 02:46:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	194164	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	191525	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	79640	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	100304	16.787	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.160%
7) Chloroethane-d5	2.805	69	95091	16.459	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.840%
11) 1,1-Dichloroethene-d2	3.916	65	19133	18.038	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.160%
21) 2-Butanone-d5	6.981	46	23854	52.196	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.400%
24) Chloroform-d	7.569	84	110308	20.319	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.280%
26) 1,2-Dichloroethane-d4	8.234	65	54848	21.901	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	8.199	84	216068	20.350	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	9.210	67	65469	21.406	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.640%
41) Toluene-d8	10.269	98	187501	18.861	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.440%
43) trans-1,3-Dichloroprop...	10.522	79	27318	21.251	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.000%
47) 2-Hexanone-d5	10.875	63	19075	48.244	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.480%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	61547	22.720	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.880%
66) 1,2-Dichlorobenzene-d4	13.810	152	64063	23.308	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.240%
Target Compounds						Qvalue
16) Methylene chloride	4.805	84	20188	5.835	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD061024\
Data File : VD079156.D
Acq On : 10 Jun 2024 14:03
Operator : RP/MD
Sample : VIBLK806
Misc : 5.00G/10ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK806

Quant Time: Jun 11 02:48:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM052924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jun 11 02:46:15 2024
Response via : Initial Calibration

